



America

CERTIFICATE

No. U8 18 03 95104 007

Holder of Certificate: Ocean Optics Inc.

4301 Metric Drive
Winter Park FL 32792
USA

Production Facility(ies): 95104

Certification Mark:



Product: Electrical equ. for measurement, control and laboratory use
Flame Spectrometer

Model(s): FLAMEX-F-XXX-YYY-ZZ and Ocean HDX
(See Attachment for Additional Model Details)

Parameters:

Rated Input Voltage:	5VDC
Rated Input Current:	2.5A
Protection Class:	Not Directly Connected to Mains.
Degree of Protection:	IPX0

Tested according to: CAN/CSA C22.2 No. 61010-1:2012
UL 61010-1:2012
EN 61010-1:2010

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 72122700-100

Date, 2018-04-02

Page 1 of 2

David B. Dorfner



Attachment to Certificate U8 18 03 95104 007

Ocean Optics Inc.
4301 Metric Drive
Winter park, FL 32792 USA



America

General Product Information:

Description of Unit:

Device is a miniature spectrometer for UV-Vis absorbance, fluorescence, Vis-NIR reflectance and IR spectroscopy.

Model numbering is FLAMEX-F-XXX-YYY-ZZ where XXX is the starting range (e.g. UV, VIS) or the full range (e.g. XR1) and YYY is the ending range (e.g. VIS, NIR) if applicable. ZZ would be the designator of sensitivity (e.g. ES) if applicable.

The device has a metal enclosure with no user serviceable parts.
The device has no connection to Mains and is powered by 5VDC that is provided via an external certified power source.

Additional Model Ocean HDX added to report, new unit uses a different light sensor.

Test Report No: 72122700-100

Date, 2018-04-02
U8 18 03 95104 007